

How to Exchange the miniReceiver Wax Guard



Step 1: Remove the click dome or click sleeve



Step 2: Take a new wax guard from the package (old packaging on the left; new packaging on the right)



Step 3: Use the end of the stick without a wax guard to remove the old wax guard from the hearing aid receiver.



Step 4: Turn the stick around and insert the new wax guard into the receiver. Dispose of the stick.



Step 5: Replace the click dome or click sleeve by inserting the receiver into the dome/sleeve until you hear a click.

